

**Altoona Water Authority
900 Chestnut Avenue
Altoona, PA, 16601**

Advertisement For Bids

Teledyne 465 Series Ozone Analyzers for Plane 9 Water Treatment Plant

The Altoona Water Authority (AWA) will receive separate and sealed bids for the above listed items at our Administrative Office located at 900 Chestnut Avenue, Altoona, PA, 16601. **Bids can be submitted until 3:00 p.m. Monday, August 21, 2017. No proposals will be accepted after this time. Sealed proposals will be publicly opened and read aloud after 1:00 p.m. Tuesday, August 22, 2017, at the 900 Chestnut Avenue location.**

Copies of the Bidding Documents are on file and open to the public for inspection at the AWA Purchasing Office located in the Administrative Building at 900 Chestnut Avenue, Altoona, PA, between the hours of 7:00 a.m. and 3:00 p.m., excepting weekends and holidays. Prospective bidders may obtain copies of the documents during these hours. Bidders may also request Bidding Documents be mailed by phoning (814) 949-2213 or faxing (814) 949-2254.

A Bid/Performance Bond of ten percent (10%) of the total bid price must accompany each proposal (NO CHECKS) to serve as a guarantee that when the contract is awarded the successful bidder will furnish the material in accordance with the specifications. In order to be received and considered, all proposals must be sealed and clearly marked “**Ozone Analyzers**”, bear the name and address of the bidder, and be addressed to:

**Altoona Water Authority
900 Chestnut Avenue
Altoona, PA, 16601
ATTN: Mr. John Swope, Purchasing Manager**

Bids must be submitted on the Form of Proposal provided or they will be rejected.

The AWA reserves the right to reject any or all bids or portions thereof, to waive any informality and to award the contract as may be deemed by the AWA to be in their best interest.

No bidder may withdraw his bid within SEVEN (7) days after the bid opening.

John H. Swope
Purchasing Manager

GENERAL

1. All bid prices will be include freight, FOB: 1923 Veterans Memorial, Altoona, PA, 16602.
2. All ozone analyzers will be Teledyne 465 series units only. No other manufacturers will be accepted.
3. 24 hour notification will be given prior to delivery. Deliveries must be made Monday through Friday, between the hours of 7 a.m. to 2:30 p.m.
4. AWA reserves the right to reject any units damaged in shipping.

SPECIFICATIONS

Model #465H NEMA high level process O3 analyzer

NEMA enclosure	1/4" Stainless steel fittings
Temperature & pressure compensation	4-20-mA or 0-5V output
Range: 0-15% by weight	Auto-zero valve for ext zero gas
Particulate filter	Flow meter
Exhaust ozone destructor	Calibration certificate

Model #465M NEMA midrange process O3 analyzer for off gas

0-1000 ppm	Calibration certificate
NEMA enclosure	Flow meter
Temperature & pressure compensation	Sample pump
Particulate filter	Internal auto-zero
1/4" Stainless steel fittings	Sample conditioner
4-20mA or 0-5 V output	Exhaust ozone destructor

Model #465L NEMA process O3 ambient (0-1000ppm) vent gas analyzer

Temperature & pressure compensation	Calibration certificate
Exhaust ozone destructor	1/4" Stainless steel fittings
Sample pump	Particulate filter
4-20mA or 0-5 v output	Flow meter
Product manual available on CD	Internal auto zero

Model #465L NEMA process O3 (0-1ppm) ambient analyzer

Temperature & pressure compensation	Calibration certificate
Exhaust ozone destructor	1/4" Stainless steel fittings
Sample pump	Particulate filter
4-20mA or 0-5V output	Flow meter
Product manuals available on CD	Internal auto zero

**Altoona Water Authority
900 Chestnut Avenue
Altoona, PA, 16601**

**Form of Proposal
Eight (8) Teledyne Series 465 Ozone Analyzers**

Pursuant to and in accordance with the legal advertisement seeking separate and sealed bid proposals to be submitted until 1:00 p.m. Monday, August 21, 2017, the undersigned agrees to furnish and deliver the equipment specified to the Altoona Water Authority's Horseshoe Curve Water Treatment Plant located at 1923 Veterans Memorial Highway, Altoona, PA, 16602. Deliveries must be made Monday through Friday, between the hours of 7:30 a.m. to 2:30 p.m.

Item #	Quantity	Description	Unit Price	Extended Price
1	2	Teledyne Model 465H NEMA high level ozone analyzer		
2	2	Teledyne Model 465M NEMA midrange Ozone Analyzer (off gas)		
3	2	Teledyne Model 465L NEMA, process O3 ambient, (0-1000PPM) ozone analyze (vent gas)		
4	2	Teledyne Model 465L NEMA, process O3 ambient, (0-1PPM) ozone analyzer		
Total Price (Total of extended prices for items 1-4)			\$	

Total Written Price: \$ _____

Bidder Name & Address _____

Bidder Contact Name _____

Authorized signature of bidder _____

Witness signature _____

Bidder contact phone # _____

Bidder contact e-mail address _____

Additional remarks or exceptions: _____
